

SPECIFICATIONS FOR WATER BUBBLER KIT

1. Bubbler quantity : 2 Nos
 - a. Purpose : Measurement of radon/thoron dissolved in water/liquid sample Kit
 - b. Material : Borosilicate glass
 - c. Bubbler capacity : Sample capacity (50-75 ml)
 - d. Head space volume < 60 ml
2. Sampler syringe : 1 No
 - a. Capacity : 300-500ml
 - b. Nozzle length : 5-10 cm
3. Sampling Bottles : 10 Nos
 - a. Capacity : 50-75 ml
 - b. Material : Borosilicate glass
 - c. Sealing : Air tight caps
4. Hard carry case for kit: 1 No